

Atomoxetine

CHROMATOGRAPHIC CONDITIONS

Column: CHIRALPAK® IC [4.6 x 250 mm, 5 µm]

Part No: 83325

Mobile phase: n-Hexane : 2-propanol : DEA (90/10/0.1, v/v)


Flowrate: 1.0 mL/min

Column Oven Temperature: 25 °C

UV Detection: 230 nm

Diluent: Mobile phase

Sample Concentration: 1.0 mg/mL



SEPARATION PERFORMANCE

t_{R_1} : 8.0	N_1 : 6630	T_1 : 1.8	
t_{R_2} : 11.7	N_2 : 8202	T_2 : 1.3	R_s : 8.2

