

Panthenol

CHROMATOGRAPHIC CONDITIONS

Column: CHIRALPAK® IG-3 [4.6 x 150 mm, 3 µm]
Part No: 87524

Mobile phase: 5 mM ammonium bicarbonate in
[(Water : Methanol 5/95, v/v)]

Flowrate: 0.7 mL/min


Column Oven Temperature: 25°C

UV Detection: 210 nm

Sample Concentration: 2.0 mg/mL

Diluent: Mobile phase

Injection Volume: 10µL



SEPARATION PERFORMANCE

tR1:	3.0	N1:	9978	T1:	1.2	
tR2:	3.6	N2:	8420	T2:	1.3	Rs: 4.3

